

[品質保証書]

本製品に使用されている ^{241}Am は以下の通り製造元で保証されています。



NRD, LLC 800-525-8076
 2937 ALT BOULEVARD 716-773-7634
 PO BOX 310 716-773-7744 FAX
 GRAND ISLAND, NY sales@nrdinc.com
 14072-0310 www.nrdinc.com

Certificate No. 035031 Customer TOKYO NUCLEAR SERVICES

Purchase Order No. RPG-06-001 Customer Dwg. No. N/A

**CERTIFICATE
 FOR
 SEALED RADIOACTIVE SOURCES**

ISO 2919 Classification designated by Code No. ISO/C32222

NRD Model No.: **A-001**

Radionuclide: **Americium 241**

Source Holder No. N/A

Activity per unit 8.5 KBq (+/- 15%) Total Activity 0.023 mCi

Chemical and Physical form: **Solid Metal Foil**

Leak test results per ISO9978 2-15-92 20,000/μci; Date 11/27/06

Total sources tested 100 Total sources shipped 100

We certify that sealed sources comply with testing requirements of ISO 2919 and the above information is correct.

We declare that we hold "IAEA Certificate of Competent Authority" No. USA-0036-S, in respect to sealed sources of Special Form Non Dispersible Radioactive Material, issued on 7-17-02, by U.S. Department of Transportation, Office of Hazardous Materials Regulations.

Date 11/27/06

Signature J. Waters

Title QC